

Olerup SSP[®] DQB1*03

Product number: 101.214-24/06 – including *Taq* pol.
101.214-24u/06u – without *Taq* pol.
Lot number: 15F
Expiry date: 2010-September-01
Number of tests: 24 test – Product No. 101.214-24
6 tests – Product No. 101.214-06
Number of wells per test: 21

CHANGES COMPARED TO THE PREVIOUS DQB1*03 LOT (Y07:

Well	5'-primer	3'-primer	Rationale
4	Moved	Moved	Primer pair moved to vial 18.
8	Moved	-	Primer moved to vial 17.
10	Moved	-	Primer moved to vial 17.
17	New	New	Primers moved from vials 8 and 10.
18	New	New	Primer pair from vial 4.
19	New	New	New primer pair for increased resolution.
20	New	New	New primer pair for the DQB1*0321 allele.
21	New	New	New primer pair for the DQB1*0322 allele.

THE NUMBER OF WELLS has been increased from 16 to 21.

ALLELE COVERAGE:

DQB1*0301 to DQB1*0322, i.e. all the currently recognized DQB1*03 alleles, give rise to unique amplification patterns; www.ebi.ac.uk/imgt/hla, 2008-July-11, release 2.22.0.

RESOLUTION IN DQB1*03 HOMOZYGOTES:

The DQB1*0301,0301, DQB1*0301,0302, DQB1*0302,0302, DQB1*0302,0303 and DQB1*0303,0303 genotypes all give rise to unique amplification patterns.

INFLUENCE ON THE INTERPRETATION OF DQB1*03 SUBTYPINGS BY NON-DQB1*03 ALLELES:

None of importance.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No suggestions have been received.