

101.212-24 – including *Taq* polymerase101.212-24u – without *Taq* polymeraseLot No.: **40K****Olerup SSP® DQB1*06**

Product number: 101.212-24 – including *Taq* polymerase
101.212-24u – without *Taq* polymerase
Lot number: 40K
Expiry date: 2012-September-01
Number of tests: 24
Number of wells per test: 30

CHANGES COMPARED TO THE PREVIOUS DQB1*06 LOT (02G).

Well	5'-primer	3'-primer	rationale
11	Added	-	Primer added for increased yield of specific PCR product.
19	Added	Added	Primer pair added for the DQB1*06:36 allele.
20	Added	-	Primer added for the DQB1*06:37 allele.
22	Added	Added	Primer pair added for the DQB1*06:38 allele.
26	Added	-	Primer added for the DQB1*06:35 allele.
27	Added	-	Primer added for the DQB1*06:40 allele.
30	New	New	New primer pair for the DQB1*06:39 allele.

THE NUMBER OF WELLS is increased from 29 to 30.

ALLELE COVERAGE:

DQB1*06:01 to DQB1*06:40 i.e. all the currently recognized DQB1*06 alleles, give rise to unique amplification patterns; www.ebi.ac.uk/imgt/hla, 2010-July-16, release 3.1.0.

RESOLUTION IN DQB1*06 HOMO- AND HETEROZYGOTES:

Very good.

INFLUENCE ON THE INTERPRETATION OF DQB1*06 SUBTYPINGS BY NON-DQB1*06 ALLELES:

None of importance.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No suggestions received.