

101.231-24/04 – including *Taq* polymerase

101.231-24u/04u – without *Taq* polymerase

Lot No.: **6R7**

Olerup SSP[®] DQA1

Product number:	101.231-24/04 – including <i>Taq</i> pol. 101.231-24u/04u – without <i>Taq</i> pol.
Lot number:	6R7
Expiry date:	2027-02-01
Number of tests:	24 tests – Product No. 101.231-24/24u 4 tests – Product No. 101.231-04/04u
Number of wells per test:	34+1

The DQA1 primer set is unchanged compared to the previous Olerup SSP[®] DQA1 (Lot No. 6N8).

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

All the DQA1 alleles, i.e. **DQA1*01:01 to 01:100**, **DQA1*02:01 to DQA1*02:29**, **DQA1*03:01 to 03:41**, **DQA1*04:01 to DQA1*04:14N**, **DQA1*05:01 to 05:70** and **DQA1*06:01 to 06:04**, recognized by the HLA Nomenclature Committee in October 2022^{1,2} will be amplified by the primers in the DQA1 typing kit, www.ebi.ac.uk/imgt/hla, 2022-October-12, release 3.50.0.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

²The DQA1 kit enables separation of DQA1 alleles as listed in the IMGT/HLA database 3.34.0.

RESOLUTION IN DQA1 HOMOZYGOTES:

Good.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.

