

101.111-24/06 – including *Taq* polymerase
101.111-24u/06u – without *Taq* polymerase

Lot No.: **90G**

Olerup SSP® DRB1*01

Product number: 101.111-24/06 – including *Taq* pol.
101.111-24u/06u – without *Taq* pol.
Lot number: 90G
Expiry date: 2012-March-01
Number of tests: 24 test – Product No. 101.111-24
6 tests – Product No. 101.111-06
Number of wells per test: 19

CHANGES COMPARED TO THE PREVIOUS DRB1*01 LOT (50F):

Well	5'-primer	3'-primer	rationale
5	-	Added	Primer added for the DRB1*0129 allele.
6	-	Added	Primer added for the DRB1*0129 and DRB1*0131 alleles.
8	Added	-	Primer added for the DRB1*0127 allele.
12	-	Added	Primer added for the DRB1*0126 and DRB1*0131 alleles.
13	-	Added	Primer added for the DRB1*0123 and DRB1*0128 alleles.
14	-	Added	Primer added for the DRB1*0123 and DRB1*0124 alleles.
15	-	Added	Primer added for the DRB1*0121 and DRB1*0128 alleles.
16	-	Added	Primer added for the DRB1*0121 and DRB1*0124 alleles.
17	New	New	New primer pair for the DRB1*0122 and DRB1*0127 alleles.
18	New	New	New primer pair for the DRB1*0125 allele.
19	New	New	New primer pair for the DRB1*0130 allele.

THE NUMBER OF WELLS is increased from 16 to 19.

ALLELE COVERAGE:

DRB1*0101 to DRB1*0131, i.e. all the currently recognized DRB1*01 alleles, give rise to unique amplification patterns; www.ebi.ac.uk/imgt/hla, 2010-January-15, release 2.28.0.

RESOLUTION IN DRB1*01 HOMOZYGOTES:

Good.

INFLUENCE ON THE INTERPRETATION OF DRB1*01 SUBTYPINGS BY NON-DRB1*01 ALLELES:

None.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No modifications made.