

DRB1*11 Release Note Page 1 of 2

101.115-24/03 – including *Taq* polymerase 101.115-24u/03u – without *Taq* polymerase

Lot No.: 8H4

Olerup SSP® DRB1*11

Product number: 101.115-24/03 – including *Taq* pol.

101.115-24u/03u – without *Tag* pol.

Lot number: 8H4

Expiry date: 2022-01-01

Number of tests: 24 tests – Product No. 101.115-24/24u

3 tests - Product No. 101.115-03/03u

Number of wells per test: 47+1

Changes compared to the previous DRB1*11 Lot (7F2):

Well	5'-primer	3'-primer	rationale
1	Added	Added	5'-primer added for the DRB1*11:01:28 allele.
			3'-primer added for the DRB1*11:04:13 allele.
3	Added	-	5'-primer added for the DRB1*11:04:13 allele.
8	Added	-	5'-primer added for the DRB1*11:04:13 allele.
12	-	Modified	3'-primer modified for improved HLA-specific amplification.
13	-	Exchanged	3'-primer exchanged for decreased tendency of primer oligomer formation.
41	Modified	-	5'-primer modified for improved HLA-specific amplification. Positive control primer exchanged for decreased tendency of primer oligomer formation.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

DRB1*11:01 to DRB1*11:233, i.e. all the currently recognized DRB1*11 alleles, will be amplified by the primers in the DRB1*11 subtyping kit¹; www.ebi.ac.uk/imgt/hla, 2018-October-18, release 3.34.0.

The DRB1*11 kit enables separation of the confirmed DRB1*11 alleles as listed in the IMGT/HLA database. An HLA allele is listed as confirmed by IMGT/HLA if it has been sequenced by more than a single laboratory or from multiple sources.

The DRB1*11 kit also enables identification of null and alternatively expressed alleles.





DRB1*11 Release Note Page 2 of 2

101.115-24/03 – including *Taq* polymerase 101.115-24u/03u – without *Taq* polymerase

Lot No.: 8H4

The following DRB1*11 alleles can be distinguished by the different sizes of the HLA-specific PCR product:

Alleles	Primer mix
DRB1*11:24:01-11:24:02,	16
11:91	
DRB1*11:61, 11:90	12
DRB1*11:97, 11:169N	34
DRB1*11:99, 11:128, 11:175	36

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page http://hla.alleles.org/alleles/deleted.html.

RESOLUTION IN DRB1*11 HOMOZYGOTES:

Good.

INFLUENCE ON THE INTERPRETATION OF DRB1*11 SUBTYPINGS BY NON-DRB1*11 ALLELES:

Most frequently encountered DRB1*03,11; DRB1*11,13 and DRB1*11,14 genotypes give rise to unique amplification patterns except for the DRB1*11:01,14:01 genotype, which gives an identical pattern as the DRB1*11:29,14:01 genotype.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.

