

101.125-24/06 – including *Taq* polymerase
101.125-24u/06u – without *Taq* polymerase

Lot No.: **35E**

Olerup SSP® DRB1*15

Product number: 101.125-24/06 – including *Taq* pol.
101.125-24u/06u – without *Taq* pol.
Lot number: 35E
Expiry date: 2009-January-01
Number of tests: 24 test – Product No. 101.125-24
6 tests – Product No. 101.125-06
Number of wells per test: 24

CHANGES COMPARED TO THE PREVIOUS DRB1*15 LOT (X37):

Well	5'-primer	3'-primer	rationale
8	New	New	Primer added for the DRB1*1527 allele.
21	New	New	New primer pair for the DRB1*1523 allele.
22	New	New	New primer pair for the DRB1*1524 allele.
23	New	New	New primer pair for the DRB1*1525 allele.
24	New	New	New primer pair for the DRB1*1526 allele.

THE NUMBER OF WELLS has been increased from 20 to 24 wells.

ALLELE COVERAGE:

DRB1*1501 to DRB1*1527, i.e. all the currently recognized DRB1*15 alleles, give rise to unique amplification patterns; www.ebi.ac.uk/imgt/hla, 2008-January-11, release 2.20.0.

RESOLUTION IN DRB1*15 HOMOZYGOTES:

Excellent.

INFLUENCE ON THE INTERPRETATION OF DRB1*15 SUBTYPINGS BY NON-DRB1*15 ALLELES:

None frequently occurring.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.