

Olerup SSP[®] DRB1*16

Product number: 101.126-12 – including *Taq* polymerase
101.126-12u – without *Taq* polymerase
Lot number: 04K
Expiry date: 2012-May-01
Number of tests: 12
Number of wells per test: 16

CHANGES COMPARED TO THE PREVIOUS DRB1*16 LOT (63F):

Well	5'-primer	3'-primer	rationale
3	-	Modified	Modified 3'primers for improved specificity.
14	New	New	New primer pair for the DRB1*16:14 allele.
15	New	New	New primer pair for the DRB1*16:15 allele.
16	New	New	New primer pair for the DRB1*16:16 allele.

THE NUMBER OF WELLS has been increased from 13 to 16.

ALLELE COVERAGE:

DRB1*16:01 to DRB1*16:16, i.e. all the currently recognized DRB1*16 alleles, give rise to unique amplification patterns; www.ebi.ac.uk/imgt/hla, 2010-April-01, release 3.0.0.

RESOLUTION IN DRB1*16 HOMOZYGOTES:

Excellent.

INFLUENCE ON THE INTERPRETATION OF DRB1*16 SUBTYPINGS BY NON-DRB1*16 ALLELES:

None frequently occurring.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.