

Olerup SSP[®] DRB1*07

Product numbers: 101.118-24 – including *Taq* polymerase
101.118-24u – without *Taq* polymerase
Lot number: 32E
Expiry date: 2010-April-01
Number of tests: 24
Number of wells per test: 12

CHANGES COMPARED TO THE PREVIOUS DRB1*07 LOT (X68):

Well	5'-primer	3'-primer	rationale
3	Removed	Removed	Primer pair for the DRB1*0712 allele moved to well 9.
8	Removed	Removed	Primer pair for the DRB1*0712 allele moved to well 9.
9	New	New	Primer pair for the DRB1*0712 allele.
10	New	New	Primer pair for the DRB1*0713 allele.
11	New	New	Primer pair for the DRB1*0714 allele.
12	New	New	Primer pair for the DRB1*0715 allele.

THE NUMBER OF WELLS has been increased from 8 to 12.

ALLELE COVERAGE:

DRB1*0701 to DRB1*0715, i.e. all the currently recognized DRB1*07 alleles, give rise to unique amplification patterns; www.ebi.ac.uk/imgt/hla, 2008-January-11, release 2.20.0.

RESOLUTION IN DRB1*07 HOMOZYGOTES:

The DRB1*0701,0701 genotype gives rise to a unique amplification pattern.

INFLUENCE ON THE INTERPRETATION OF DRB1*07 SUBTYPINGS BY NON-DRB1*07 ALLELES:

None.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No suggestions received.