

## **Olerup SSP<sup>®</sup> DRB1\*14:01/14:54**

<b>Product number:</b>	101.813-12 – including <i>Taq</i> polymerase 101.813-12u – without <i>Taq</i> polymerase
<b>Lot number:</b>	45N
<b>Expiry date:</b>	2014-August-01
<b>Number of tests:</b>	12
<b>Number of wells per test:</b>	2

### **CHANGES COMPARED TO THE PREVIOUS DRB1\*14:01/14:54 Lot(12L).**

The DRB1\*14:01/14:54 primer set is unchanged compared to the previous lot.

**THE NUMBER OF WELLS** is unchanged.

### **ALLELE COVERAGE:**

The DRB1\*14:01 and DRB1\*14.54 alleles will give rise to unique amplification patterns by the primers in the DRB1\*14:01/14:54 kit; [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla), 2012-January-12, release 3.7.0.

### **RESOLUTION IN DRB1\*14:01/14:54 HOMOZYGOTES:**

Excellent.

### **INFLUENCE ON THE INTERPRETATION OF DRB1\*14:01/14:54 SUBTYPINGS BY NON-DRB1\*14:01/14:54 ALLELES:**

None of importance.

### **MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:**

No suggestions received.