■LERUP SSP* DRB1*14:01/14:54

Release Note

Page 1 of 1

101.813-12 – including *Taq* polymerase 101.813-12u – without *Taq* polymerase

Lot No.: **0G7**

Olerup SSP® DRB1*14:01/14:54

Product number: 101.813-12 – including *Taq* polymerase

101.813-12u - without *Tag* polymerase

Lot number: 0G7

Expiry date: 2020-08-01

Number of tests: 12 Number of wells per test: 2+1

The DRB1*14:01/14:54 primer set is unchanged compared to the previous *Olerup* SSP® DRB1*14:01/14:54 (Lot No. 7D6).

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

The DRB1*14:01 and DRB1*14:54 alleles will give rise to unique amplification patterns by the primers in the DRB1*14:01/14:54 kit;¹ www.ebi.ac.uk/imgt/hla, 2017-October-27, release 3.30.0.

RESOLUTION IN DRB1*14:01/14:54 HOMOZYGOTES:

Excellent.

INFLUENCE ON THE INTERPRETATION OF DRB1*14:01/14:54 SUBTYPINGS BY NON-DRB1*14:01/14:54 ALLELES:

None of importance.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No suggestions received.



¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page http://hla.alleles.org/alleles/deleted.html.