

Olerup SSP[®] DRB4*01:03:01:02N

Product number:	101.811-12 – including <i>Taq</i> polymerase 101.811-12u – without <i>Taq</i> polymerase
Lot number:	6F3
Expiry date:	2021-10-01
Number of tests:	12
Number of wells per test:	2+1

The DRB4*01:03:01:02N primer set is unchanged compared to the previous *Olerup SSP[®] DRB4*01:03:01:02N (Lot No. 6E4)*.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

The DRB4*01:03:01:02N allele will give rise to a unique amplification pattern by the primers in the DRB4*01:03:01:02N kit; www.ebi.ac.uk/imgt/hla, 2017-April-13, release 3.28.0¹.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

RESOLUTION IN DRB4*01:03:01:02N HOMOZYGOTES:

Excellent.

INFLUENCE ON THE INTERPRETATION OF DRB4*01:03:01:02N SUBTYPINGS BY NON-DRB4*01:03:01:02N ALLELES:

None of importance.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.

Changes in revision R01 compared to R00:

1. The expiration date has been altered due to extension of shelf-life.