

## **Olerup SSP<sup>®</sup> DRB5**

**Product number:** 101.123-24/06 – including *Taq* pol.  
101.123-24u/06u – without *Taq* pol.  
**Lot number:** 28E  
**Expiry date:** 2010-March-01  
**Number of tests:** 24 test – Product No. 101.123-24  
6 tests – Product No. 101.123-06  
**Number of wells per test:** 16

### **CHANGES COMPARED TO THE PREVIOUS DRB5 LOT (X51):**

The DRB5 primer set is unchanged.

**THE NUMBER OF WELLS** is unchanged.

### **ALLELE COVERAGE:**

DRB5\*0101 to DRB5\*0113 and DRB5\*0202 to DRB5\*0205, i.e. all the currently recognized DRB5 alleles, give rise to unique amplification patterns; [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla), 2008-January-11, release 2.20.0.

### **RESOLUTION IN DRB5 HOMOZYGOTES:**

Excellent.

### **INFLUENCE ON THE INTERPRETATION OF DRB5 SUBTYPINGS BY NON-DRB5 ALLELES:**

None.

### **MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:**

No comments received.