OLERUPSSP[®] DRB5 Release Note 101.123-24/06 – including *Taq* polymerase 101.123-24u/06u – without *Taq* polymerase Lot No.: **25V**

Olerup SSP[®] DRB5

Product number:	101.123-24/06 – including <i>Taq</i> pol.
	101.123-24u/06u – without <i>Taq</i> pol.
Lot number:	25V
Expiry date:	2016-July-01
Number of tests:	24 test – Product No. 101.123-24/24u
	6 tests – Product No. 101.123-06/06u
Number of wells per test:	15+1

CHANGES COMPARED TO THE PREVIOUS DRB5 LOT (49R):

The primers of the wells detailed below have been exchanged, added or modified compared to the previous lot.

Well	5'-primer	3'-primer	rationale
7		Added	3'-primer added from well 16.
16	Moved	Moved	Primer pair moved to well 7, Negative Control.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

DRB5*01:01:01 to DRB5*01:14 and DRB5*02:02 to DRB5*02:06, i.e. all the currently recognized DRB5 alleles, give rise to unique amplification patterns, <u>www.ebi.ac.uk/imgt/hla</u>, 2013-October-11, release 3.14.0.

RESOLUTION IN DRB5 HOMOZYGOTES:

Good.

INFLUENCE ON THE INTERPRETATION OF DRB5 SUBTYPINGS BY NON-DRB5 ALLELES: None frequently occurring.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.

