

101.123-24/06 – including *Taq* polymerase

101.123-24u/06u – without *Taq* polymerase

Lot No.: **49X**

Olerup SSP® DRB5

Product number: 101.123-24/06 – including *Taq* pol.
101.123-24u/06u – without *Taq* pol.

Lot number: 49X

Expiry date: 2017-May-01

Number of tests: 24 test – Product No. 101.123-24/24u
6 tests – Product No. 101.123-06/06u

Number of wells per test: 15+1

CHANGES COMPARED TO THE PREVIOUS DRB5 LOT (25V):

The primers of the wells detailed below have been exchanged, added or modified compared to the previous lot.

| Well | 5'-primer | 3'-primer | rationale |
|------|-----------|-----------|--------------------------|
| 16 | - | - | Updated negative control |

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

DRB5*01:01:01 to DRB5*01:15 and DRB5*02:02 to DRB5*02:06, i.e. all the currently recognized DRB5 alleles¹, give rise to unique amplification patterns, www.ebi.ac.uk/imgt/hla, 2014-July-25, release 3.17.0.

¹ Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

RESOLUTION IN DRB5 HOMOZYGOTES:

Good.

INFLUENCE ON THE INTERPRETATION OF DRB5 SUBTYPINGS BY NON-DRB5 ALLELES:

None frequently occurring.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.