

101.123-24/06 – including *Taq* polymerase

101.123-24u/06u – without *Taq* polymerase

Lot No.: **79Y**

Olerup SSP[®] DRB5

Product number: 101.123-24/06 – including *Taq* pol.
 101.123-24u/06u – without *Taq* pol.

Lot number: 79Y

Expiry date: 2018-February-01

Number of tests: 24 test – Product No. 101.123-24/24u
 6 tests – Product No. 101.123-06/06u

Number of wells per test: 15+1

CHANGES COMPARED TO THE PREVIOUS DRB5 LOT (49X):

Well	5'-primer	3'-primer	rationale
4	-	-	Exchanged positive control primer pair for decreased tendency of primer oligomer formation.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

DRB5*01:01:01 to DRB5*01:15 and DRB5*02:02 to DRB5*02:06, i.e. all the currently recognized DRB5 alleles¹, give rise to unique amplification patterns, www.ebi.ac.uk/imgt/hla, 2015-August-11, release 3.21.1.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

RESOLUTION IN DRB5 HOMOZYGOTES:

Good.

INFLUENCE ON THE INTERPRETATION OF DRB5 SUBTYPINGS BY NON-DRB5 ALLELES:

None frequently occurring.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.