

101.123-24/06 – including *Taq* polymerase

101.123-24u/06u – without *Taq* polymerase

Lot No.: **1V4**

Olerup SSP[®] DRB5

Product number:	101.123-24/06 – including <i>Taq</i> pol. 101.123-24u/06u – without <i>Taq</i> pol.
Lot number:	1V4
Expiry date:	2029-02-01
Number of tests:	24 test – Product No. 101.123-24/24u 6 tests – Product No. 101.123-06/06u
Number of wells per test:	21+1

CHANGES COMPARED TO THE PREVIOUS DRB5 LOT (0V1):

The DRB5 primer set is unchanged compared to the previous *Olerup SSP[®] DRB5 (Lot No. 0V1)*.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

DRB5*01:01:01 to DRB5*01:142 and DRB5*02:02 to DRB5*02:39, i.e. all the currently recognized DRB5 alleles¹, will be amplified by the primers in the DRB5 subtyping kit, www.ebi.ac.uk/imgt/hla, 2024-October-09, release 3.58.0².

The DRB5 kit enables separation of the confirmed DRB5 alleles as listed in the IMGT/HLA database 3.28.0. An HLA allele is listed as confirmed by IMGT/HLA if it has been sequenced by more than a single laboratory or from multiple sources.

The DRB5 kit also enables identification of many null and alternatively expressed alleles.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

²The primers in this kit do not amplify the DRB5*01:131N allele.

RESOLUTION IN DRB5 HOMOZYGOTES:

Good.

INFLUENCE ON THE INTERPRETATION OF DRB5 SUBTYPINGS BY NON-DRB5 ALLELES:

None frequently occurring.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.



0197