

## **Olerup SSP<sup>®</sup> HLA-A\*23**

**Product number:** 101.421-06 – including *Taq* polymerase  
101.421-06u – without *Taq* polymerase  
**Lot number:** 31G  
**Expiry date:** 2011-October-01  
**Number of tests:** 6  
**Number of wells per test:** 24

### **CHANGES COMPARED TO THE PREVIOUS HLA-A\*23 LOT (33F):**

Well	5'-primer	3'-primer	rationale
23	New	New	New primer pair for the A*2320 allele.
24	New	New	New primer pair for the A*2321 allele.

**THE NUMBER OF WELLS** has been increased from 22 to 24 wells.

#### **ALLELE COVERAGE:**

A\*2301 to A\*2321, i.e. all the currently recognized HLA-A\*23 alleles, give rise to unique amplification patterns; [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla), 2009-July-17, release 2.26.0.

#### **RESOLUTION IN HLA-A\*23 HOMO- AND HETEROZYGOTES:**

The A\*2301,2301 and A\*2301,2317 genotypes give rise to identical amplification patterns.

#### **INFLUENCE ON THE INTERPRETATION OF HLA-A\*23 SUBTYPINGS BY NON-HLA-A\*23 ALLELES:**

None frequently occurring.

#### **MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:**

No comments received.