■LERUP SSP®

HLA-A*23 Release Note Page 1 of 2

101.421-06 – including *Taq* polymerase 101.421-06u – without *Taq* polymerase

Lot No.: **0G9**

Olerup SSP® HLA-A*23

Product number: 101.421-06 – including *Taq* polymerase

101.421-06u - without *Tag* polymerase

Lot number: 0G9

Expiry date: 2022-03-01

Number of tests: 6
Number of wells per test: 35+1

CHANGES COMPARED TO THE PREVIOUS HLA-A*23 LOT (7E3):

Well	5'-primer	3'-primer	rationale
23	Modified	-	5'-primer modified for improved HLA-specific
			amplification.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

A*23:01 to A*23.83, i.e. all the currently recognized HLA-A*23 alleles, will be amplified by the primers in the HLA-A*23 subtyping kit¹; www.ebi.ac.uk/imgt/hla, 2017-October-27, release 3.30.0.

The HLA-A*23 kit enables separation of the confirmed HLA-A*23 alleles as listed in the IMGT/HLA database. An HLA allele is listed as confirmed by IMGT/HLA if it has been sequenced by more than a single laboratory or from multiple sources.

The HLA-A*23 kit also enables identification of polymorphisms in exons outside of the region encoding the peptide binding domain and of null and alternatively expressed alleles.

The following HLA-A*23 alleles can be distinguished by the different sizes of the HLA-specific PCR product:

Alleles	Primer mix	Alleles	Primer mix
A*23:08N, 23:22	7	A*23:13, 23:33	17
A*23:09, 23:26	8	A*23:16, 23:29	20
A*23:10, 23:23	10	A*23:41, 23:42	25

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page http://hla.alleles.org/alleles/deleted.html.

RESOLUTION IN HLA-A*23 HOMO- AND HETEROZYGOTES: Good.

INFLUENCE ON THE INTERPRETATION OF HLA-A*23 SUBTYPINGS BY NON-HLA-A*23 ALLELES:

None frequently occurring.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.





HLA-A*23 Release Note

101.421-06 – including *Taq* polymerase 101.421-06u – without *Taq* polymerase

Lot No.: **0G9**

Changes in revision R01 compared to R00:

1. The expiration date has been altered due to extension of shelf-life.



Page 2 of 2