

101.843-12 – including *Taq* polymerase

101.843-12u – without *Taq* polymerase

Lot No.: **8F8**

Olerup SSP[®] HLA-A*23 Add-on

Product number:	101.843-12 – including <i>Taq</i> polymerase 101.843-12u – without <i>Taq</i> polymerase
Lot number:	8F8
Expiry date:	2020-05-01
Number of tests:	12
Number of wells per test:	3+1

The HLA-A*23 Add-on primer set is unchanged compared to the previous *Olerup SSP[®] HLA-A*23 Add-on (Lot No. 6E6)*.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

The HLA-A*23:01 and 23:17 alleles give different patterns in the HLA-A*23 Add-on kit; www.ebi.ac.uk/imgt/hla, 2017-August-10, release 3.29.0.

RESOLUTION IN HOMO- AND HETEROZYGOTES:

Excellent.

INFLUENCE ON THE INTERPRETATION OF HLA-A*23 ADD-ON SUBTYPINGS BY NON-HLA-A*23 ALLELES:

None of importance.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.