■LERUP SSP*

HLA-A*24 Add-on Release Note Page 1 of 1

101.846-12 – including *Taq* polymerase 101.846-12u – without *Taq* polymerase

Lot No.: 3F2

Olerup SSP® HLA-A*24 Add-on

Product number: 101.846-12 – including *Taq* polymerase

101.846-12u - without *Tag* polymerase

Lot number: 3F2

Expiry date: 2020-01-01

Number of tests: 12 Number of wells per test: 3+1

The HLA-A*24 Add-on primer set is unchanged compared to the previous *Olerup* SSP® HLA-A*24 Add-on (Lot No. 8E0).

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

The HLA-A*24:09N, A*24:11N and A*24:02:01G alleles give different patterns in the HLA-A*24 Add-on kit¹; www.ebi.ac.uk/imgt/hla, 2016-October-14, release 3.26.0.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page http://hla.alleles.org/alleles/deleted.html.

RESOLUTION IN HOMO- AND HETEROZYGOTES:

Excellent.

INFLUENCE ON THE INTERPRETATION OF HLA-A*24 ADD-ON SUBTYPINGS BY NON-HLA-A*24 ALLELES:

None of importance.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.

