

Olerup SSP® HLA-A*24:09N

Product number: 101.841-12 – including *Taq* polymerase
 101.841-12u – without *Taq* polymerase

Lot number: 0D6

Expiry date: 2018-April-01

Number of tests: 12

Number of wells per test: 2+1

CHANGES COMPARED TO THE PREVIOUS HLA-A*24:09N LOT (48V):

Well	5'-primer	3'-primer	rationale
1	Exchanged	Exchanged	Primer pair exchanged for increased yield of A*24:09N allele.
3	-	-	Updated negative control.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

The HLA-A*24:09N allele will give rise to a unique amplification pattern by the primers in the HLA-A*24:09N kit; www.ebi.ac.uk/imgt/hla, 2015-August-11¹, release 3.21.1.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

RESOLUTION IN HLA-A*24:09N HOMO- AND HETEROZYGOTES:

Excellent.

INFLUENCE ON THE INTERPRETATION OF HLA-A*24:09N SUBTYPINGS BY NON-HLA-A*24:09N ALLELES:

None.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No suggestions received.