

## **Olerup SSP® HLA-A\*24:09N**

<b>Product number:</b>	<b>101.841-12 – including <i>Taq</i> polymerase</b> <b>101.841-12u – without <i>Taq</i> polymerase</b>
<b>Lot number:</b>	<b>2E4</b>
<b>Expiry date:</b>	<b>2019-05-01</b>
<b>Number of tests:</b>	<b>12</b>
<b>Number of wells per test:</b>	<b>2+1</b>

**CHANGES COMPARED TO THE PREVIOUS HLA-A\*24:09N LOT (0D6):**

<b>Well</b>	<b>5'-primer</b>	<b>3'-primer</b>	<b>rationale</b>
1	Modified	-	5'-primer modified for improved HLA-specific amplification.

**THE NUMBER OF WELLS** is unchanged.

**ALLELE COVERAGE:**

The HLA-A\*24:09N allele will give rise to a unique amplification pattern by the primers in the HLA-A\*24:09N kit; [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla), 2016-July-14<sup>1</sup>, release 3.25.0.

<sup>1</sup>Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

**RESOLUTION IN HLA-A\*24:09N HOMO- AND HETEROZYGOTES:**

Excellent.

**INFLUENCE ON THE INTERPRETATION OF HLA-A\*24:09N SUBTYPINGS BY NON-HLA-A\*24:09N ALLELES:**

None.

**MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:**

No suggestions received.