

**101.841-12 – including *Taq* polymerase**

**101.841-12u – without *Taq* polymerase**

**Lot No.: 2K2**

## **Olerup SSP® HLA-A\*24:09N**

<b>Product number:</b>	<b>101.841-12 – including <i>Taq</i> polymerase</b> <b>101.841-12u – without <i>Taq</i> polymerase</b>
<b>Lot number:</b>	<b>2K2</b>
<b>Expiry date:</b>	<b>2023-09-01</b>
<b>Number of tests:</b>	<b>12</b>
<b>Number of wells per test:</b>	<b>3+1</b>

The HLA-A\*24:09N primer set is unchanged compared to the previous Olerup SSP® HLA-A\*24:09N (Lot No. 8H6).

**THE NUMBER OF WELLS** is 4 in an 8 well plate.

### **ALLELE COVERAGE:**

The HLA-A\*24:09N allele will give rise to a unique amplification pattern by the primers in the HLA-A\*24:09N kit; [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla), 2019-April-17<sup>1</sup>, release 3.36.0.

<sup>1</sup>Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

### **RESOLUTION IN HLA-A\*24:09N HOMO- AND HETEROZYGOTES:**

Excellent.

### **INFLUENCE ON THE INTERPRETATION OF HLA-A\*24:09N SUBTYPINGS BY NON-HLA-A\*24:09N ALLELES:**

None.

### **MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:**

No suggestions received.