

HLA-A\*33  
101.432-12 – licensed for PCR  
101.432-12u – not licensed for PCR  
Lot No.: X90

1

## Release Note - Olerup SSP™ HLA-A\*33 – Lot No. X90

Product number: 101.432-12 – licensed for PCR  
101.432-12u – not licensed for PCR  
Lot number: X90  
Expiry date: 2009-May-01  
Number of tests: 12  
Number of tubes per test: 11

### CHANGES COMPARED TO THE PREVIOUS HLA-A\*33 LOT (LV3):

Tube	5'-primer	3'-primer	rationale
5	Removed	Removed	One primer pair moved to vial 9.
9	New	New	Primer pair from vial 5.
10	New	New	New primer pair for A*3308 allele.
11	New	New	New primer pair for A*3310 allele.

**THE NUMBER OF WELLS** has been increased from 8 to 11.

### ALLELE COVERAGE:

A\*3301 to A\*3310, i.e. all the currently recognized HLA-A\*33 alleles, give rise to unique amplification patterns; [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla), 2007-April-12, release 2.17.0.

### RESOLUTION IN HLA-A\*33 HOMO- AND HETEROZYGOTES:

The A\*3301,3301, A\*3301,3303 and A\*3303,3303 genotypes give rise to unique amplification patterns.

### INFLUENCE ON THE INTERPRETATION OF HLA-A\*33 SUBTYPINGS BY NON-HLA-A\*33 ALLELES:

None frequently occurring.

### MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.