

Olerup SSP[®] HLA-A*34

Product number: 101.425-06 – including *Taq* polymerase
101.425-06u – without *Taq* polymerase
Lot number: 58F
Expiry date: 2011-February-01
Number of tests: 6
Number of wells per test: 8

CHANGES COMPARED TO THE PREVIOUS HLA-A*34 LOT (52E):

The HLA-A*34 primer set is unchanged.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

A*3401 to A*3408, i.e. all the currently recognized HLA-A*34 alleles, give rise to unique amplification patterns; www.ebi.ac.uk/imgt/hla, 2009-January-16, release 2.24.0.

RESOLUTION IN HLA-A*34 HOMO- AND HETEROZYGOTES:

Good.

INFLUENCE ON THE INTERPRETATION OF HLA-A*34 SUBTYPINGS BY NON-HLA-A*34 ALLELES:

None of importance.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No modifications made.