

Lot No.: **73M**

## **Olerup SSP<sup>®</sup> HLA-A\*34**

**Product number:** 101.425-06 – including *Taq* polymerase  
101.425-06u – without *Taq* polymerase  
**Lot number:** 73M  
**Expiry date:** 2014-April-01  
**Number of tests:** 6  
**Number of wells per test:** 8

### **CHANGES COMPARED TO THE PREVIOUS HLA-A\*34 LOT (23K):**

<b>Well</b>	<b>5'-primer</b>	<b>3'-primer</b>	<b>rationale</b>
8	Added	Added	Added primer pair for the A*34:09 allele, exchanged positive control primer pair.

**THE NUMBER OF WELLS** is unchanged.

### **ALLELE COVERAGE:**

A\*34:01 to A\*34:09, i.e. all the currently recognized HLA-A\*34 alleles, give rise to unique amplification patterns; [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla), 2011-July-14, release 3.5.0.

### **RESOLUTION IN HLA-A\*34 HOMO- AND HETEROZYGOTES:**

Good.

### **INFLUENCE ON THE INTERPRETATION OF HLA-A\*34 SUBTYPINGS BY NON-HLA-A\*34 ALLELES:**

None of importance.

### **MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:**

No comments received.