■LERUP SSP°

HLA-A*34 Release Note Page 1 of 1

101.425-06 – including *Taq* polymerase 101.425-06u – without *Taq* polymerase

Lot No.: 1F0

Olerup SSP® HLA-A*34

Product number: 101.425-06 – including *Taq* polymerase

101.425-06u – without *Taq* polymerase

Lot number: 1F0

Expiry date: 2019-10-01

Number of tests: 6 Number of wells per test: 11+1

CHANGES COMPARED TO THE PREVIOUS HLA-A*34 LOT (87Y):

Well	5'-primer	3'-primer	rationale
8	-	-	Strength of control band has been optimized for
			increased yield.

THE NUMBER OF WELLS is unchanged

ALLELE COVERAGE:

A*34:01 to A*34:17, i.e. all the currently recognized HLA-A*34 alleles, will be amplified by the primers in the HLA-A*34 subtyping kit; www.ebi.ac.uk/imgt/hla, 2016-October-14¹, release 3.26.0.

RESOLUTION IN HLA-A*34 HOMO- AND HETEROZYGOTES: Good.

INFLUENCE ON THE INTERPRETATION OF HLA-A*34 SUBTYPINGS BY NON-HLA-A*34 ALLELES:

None of importance.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.



¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page http://hla.alleles.org/alleles/deleted.html.