

HLA-A*34
101.425-06 – including *Taq* polymerase
101.425-06u – without *Taq* polymerase
Lot No.: 2R2

Olerup SSP® HLA-A*34

Product number: 101.425-06 – including *Taq* polymerase
101.425-06u – without *Taq* polymerase
Lot number: 2R2
Expiry date: 2026-08-01
Number of tests: 6
Number of wells per test: 11+1

The HLA-A*34 primer set is unchanged compared to the previous *Olerup SSP*® HLA-A*34 (Lot No. 1H6).

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

A*34:01 to A*34:30N, i.e. all the currently recognized HLA-A*34 alleles, will be amplified by the primers in the HLA-A*34 subtyping kit; www.ebi.ac.uk/imgt/hla, 2022-July-12¹, release 3.49.0.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

RESOLUTION IN HLA-A*34 HOMO- AND HETEROZYGOTES:

Good.

INFLUENCE ON THE INTERPRETATION OF HLA-A*34 SUBTYPINGS BY NON-HLA-A*34 ALLELES:

None of importance.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.



0197

For *In Vitro* Diagnostic Use
MA125 v01 SSP Release Note Template
Date: August 2022, Rev. No: 00