

## **Olerup SSP<sup>®</sup> HLA-A\*74**

**Product number:** 101.433-06 – including *Taq* polymerase  
101.433-06u – without *Taq* polymerase  
**Lot number:** 87K  
**Expiry date:** 2013-July-01  
**Number of tests:** 6  
**Number of wells per test:** 15

### **CHANGES COMPARED TO THE PREVIOUS HLA-A\*74 LOT (26G):**

The HLA-A\*74 primer set is unchanged compared to the previous lot.

**THE NUMBER OF WELLS** is unchanged.

### **ALLELE COVERAGE:**

A\*74:01 to A\*74:14N, i.e. all the currently recognized HLA-A\*74 alleles, give rise to unique amplification patterns; [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla), 2010-October-15, release 3.2.0.

### **RESOLUTION IN HLA-A\*74 HOMO- AND HETEROZYGOTES:**

The A\*74:01,74:01 genotype gives rise to a unique amplification pattern.

### **INFLUENCE ON THE INTERPRETATION OF HLA-A\*74 SUBTYPINGS BY NON-HLA-A\*74 ALLELES:**

None frequently occurring.

### **MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:**

No comments received.