

Olerup SSP® HLA-A*80

Product number:	101.434-06 – including <i>Taq</i> polymerase 101.434-06u – without <i>Taq</i> polymerase
Lot number:	5S1
Expiry date:	2027-10-01
Number of tests:	6
Number of wells per test:	4+1

The HLA-A*80 primer set is unchanged compared to the previous *Olerup SSP*® HLA-A*80 (Lot No. 1L8).

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

A*80:01 and A*80:09N, i.e. all the currently recognized HLA-A*80 alleles, will be amplified by the primers in the HLA-A*80 subtyping kit¹; www.ebi.ac.uk/imgt/hla, 2023-July-12, release 3.53.0.

The HLA-A*80 kit enables separation of the confirmed HLA-A*80 alleles as listed in the IMGT/HLA database 3.28.0. An HLA allele is listed as confirmed by IMGT/HLA if it has been sequenced by more than a single laboratory or from multiple sources.

The HLA-A*80 kit also enables identification of many null and alternatively expressed alleles.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

RESOLUTION IN HLA-A*80 HOMO- AND HETEROZYGOTES:

Excellent.

INFLUENCE ON THE INTERPRETATION OF HLA-A*80 SUBTYPINGS BY NON-HLA-A*80 ALLELES:

None of importance.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No suggestions received.



0197