OLERUP SSP

HLA-B*41 Release Note Page 1 of 2

101.542-06 – including *Taq* polymerase 101.542-06u – without *Taq* polymerase

Lot No.: 8R8

Olerup SSP® HLA-B*41

Product number: 101.542-06 – including *Taq* polymerase

101.542-06u - without *Tag* polymerase

Lot number: 8R8

Expiry date: 2027-05-01

Number of tests: 6 Number of wells per test: 23+1

CHANGES COMPARED TO THE PREVIOUS HLA-B*41 LOT (0R3).

Well	5'-primer	3'-primer	rationale
16	-	Added	3'-primer added for the B*41:18 allele.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

B*41:01 to B*41:79 i.e. all the currently recognized HLA-B*41 alleles, will be amplified by the primers in the HLA-B*41 SSP kit ^{1,2}; <u>www.ebi.ac.uk/imgt/hla</u>, 2023-January-12, release 3.51.0.

The HLA-B*41 kit enables separation of the confirmed HLA-B*41 alleles as listed in the IMGT/HLA database 3.27.0. An HLA allele is listed as confirmed by IMGT/HLA if it has been sequenced by more than a single laboratory or from multiple sources.

The HLA-B*41 kit also enables identification of many null and alternatively expressed alleles.

The following HLA-B*41 alleles can be distinguished by the different sizes of the HLA-specific PCR product:

Alleles	Primer mix
B*41:04, 41:36	4
R*41·10 41·45N	3

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page http://hla.alleles.org/alleles/deleted.html.

Alleles

B*41:08, B*40:136, 40:231, 40:279 B*41:23, B*42:27 B*41:46-41:47, B*44:166 B*41:62, B*40:02:20



For *In Vitro* Diagnostic Use MA125 v01 SSP Release Note Template Date: June 2023, Rev. No: 00

²The following alleles give rise to identical amplification patterns with the HLA-B*41 high resolution kit. These alleles can be distinguished by the HLA-B low resolution and/or the respective kits.

•LERUP SSP*

HLA-B*41 Release Note Page 2 of 2

101.542-06 – including *Taq* polymerase 101.542-06u – without *Taq* polymerase

Lot No.: 8R8

RESOLUTION IN HLA-B*41 HOMO- AND HETEROZYGOTES:

Good.

INFLUENCE ON THE INTERPRETATION OF HLA-B*41 SUBTYPINGS BY NON-HLA-B*41

None frequently occurring.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.