

## **Olerup SSP<sup>®</sup> HLA-B\*54**

**Product number:** 101.569-06 – including *Taq* polymerase  
101.569-06u – without *Taq* polymerase  
**Lot number:** 95N  
**Expiry date:** 2014-December-01  
**Number of tests:** 6  
**Number of wells per test:** 20

### **CHANGES COMPARED TO THE PREVIOUS HLA-B\*54 LOT (30M):**

The HLA-B\*54 primer set is unchanged compared to the previous lot.

**THE NUMBER OF WELLS** is unchanged.

### **ALLELE COVERAGE:**

B\*54:01 to B\*54:26, i.e. all the currently recognized HLA-B\*54 alleles, will be amplified by the primers in the HLA-B\*54 subtyping kit; [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla), 2012-April-12, release 3.8.0.

The HLA-B\*54 kit enables separation of the confirmed HLA-B\*54 alleles as listed in the IMGT/HLA database. An HLA allele is listed as confirmed by IMGT/HLA if it has been sequenced by more than a single laboratory or from multiple sources.

The HLA-B\*54 kit also enables identification of polymorphisms in exons outside of the region encoding the peptide binding domain and of null and alternatively expressed alleles.

### **RESOLUTION IN HLA-B\*54 HOMO- AND HETEROZYGOTES:**

Excellent.

### **INFLUENCE ON THE INTERPRETATION OF HLA-B\*54 SUBTYPINGS BY NON-HLA-B\*54 ALLELES:**

None of importance.

### **MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:**

No comments received.

.