

Olerup SSP[®] HLA-B*57

Product number:	101.567-12 – including <i>Taq</i> polymerase 101.567-12u – without <i>Taq</i> polymerase
Lot number:	3H0
Expiry date:	2021-07-01
Number of tests:	12
Number of wells per test:	31+1

CHANGES COMPARED TO THE PREVIOUS HLA-B*57 LOT (7E2):

Well	5'-primer	3'-primer	rationale
21	Added	Modified	5'-primer added for the B*57:98Q allele, modified 3'-primer for improved HLA-specific amplification.
22	Added	-	5'-primer added for the B*57:98Q allele.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

B*57:01 to B*57:110, i.e. all the currently recognized HLA-B*57 alleles, will be amplified by the primers in the HLA-B*57 subtyping kit; www.ebi.ac.uk/imgt/hla, 2018-October-18, release 3.34.0¹.

The HLA-B*57 kit enables separation of the confirmed HLA-B*57 alleles as listed in the IMGT/HLA database 3.25.0. An HLA allele is listed as confirmed by IMGT/HLA if it has been sequenced by more than a single laboratory or from multiple sources.

The HLA-B*57 kit also enables identification of null and alternatively expressed alleles.

The following HLA-B*57 alleles can be distinguished by the different sizes of the HLA-specific PCR product:

Alleles	Primer mix
B*57:15, 57:37, 57:55	7
B*57:25, 57:41	19

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

RESOLUTION IN HLA-B*57 HOMO- AND HETEROZYGOTES:

Good.

101.567-12 – including *Taq* polymerase

101.567-12u – without *Taq* polymerase

Lot No.: 3H0

INFLUENCE ON THE INTERPRETATION OF HLA-B*57 SUBTYPINGS BY NON-HLA-B*57 ALLELES:

None frequently occurring.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.