•LERUP SSP\*

HLA-B\*57:01 Release Note Page 1 of 1

101.572-12 – including *Taq* polymerase 101.572-12u – without *Taq* polymerase

Lot No.: 3R3

## Olerup SSP® HLA-B\*57:01

Product number: 101.572-12 – including *Taq* polymerase

**101.572-12u – without** *Taq* **polymerase** 

Lot number: 3R3

Expiry date: 2026-08-01

Number of tests: 12 Number of wells per test: 23+1

### CHANGES COMPARED TO THE PREVIOUS HLA-B\*57:01 LOT (8L3):

Well	5'-primer	3'-primer	rationale
1	Added	-	3'-primer added for the B*57:01:45 allele.

THE NUMBER OF WELLS is unchanged.

#### **ALLELE COVERAGE:**

B\*57:01:01 to B\*57:01:46, i.e. all the currently recognized HLA-B\*57:01 alleles, gives rise to a unique amplification pattern<sup>1</sup>; <a href="www.ebi.ac.uk/imgt/hla">www.ebi.ac.uk/imgt/hla</a>, 2022-July-12, release 3.49.0.

<sup>1</sup>Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <a href="http://hla.alleles.org/alleles/deleted.html">http://hla.alleles.org/alleles/deleted.html</a>.

# **RESOLUTION IN HLA-B\*57:01 HOMO- AND HETEROZYGOTES:** Good.

INFLUENCE ON THE INTERPRETATION OF HLA-B\*57:01 SUBTYPINGS BY NON-HLA-B\*57 ALLELES:

None of importance.

#### MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No suggestions received.

Changes in revision R01 compared to R00:

1. The lot number in the header has been corrected.



For *In Vitro* Diagnostic Use MA125 v01 SSP Release Note Template Date: November 2023, Rev. No: 01