

HLA-B\*57:01  
101.572-12 – including *Taq* polymerase  
101.572-12u – without *Taq* polymerase  
Lot No.: **0S7**

## Olerup SSP<sup>®</sup> HLA-B\*57:01

Product number:	101.572-12 – including <i>Taq</i> polymerase 101.572-12u – without <i>Taq</i> polymerase
Lot number:	0S7
Expiry date:	2027-05-01
Number of tests:	12
Number of wells per test:	23+1

### CHANGES COMPARED TO THE PREVIOUS HLA-B\*57:01 LOT (3R3):

The HLA-B\*57:01 primer set is unchanged compared to the previous lot.

THE NUMBER OF WELLS is unchanged.

### ALLELE COVERAGE:

B\*57:01:01 to B\*57:01:46, i.e. all the currently recognized HLA-B\*57:01 alleles, gives rise to a unique amplification pattern<sup>1</sup>; [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla), 2023-January-12, release 3.51.0.

<sup>1</sup>Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

### RESOLUTION IN HLA-B\*57:01 HOMO- AND HETEROZYGOTES:

Good.

### INFLUENCE ON THE INTERPRETATION OF HLA-B\*57:01 SUBTYPINGS BY NON-HLA-B\*57 ALLELES:

None of importance.

### MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No suggestions received.



0197

For *In Vitro* Diagnostic Use  
MA125 v01 SSP Release Note Template  
Date: June 2023, Rev. No: 00