

Olerup SSP® HLA-B*07 Add-on

Product number:	101.852-12 – including <i>Taq</i> polymerase 101.852-12u – without <i>Taq</i> polymerase
Lot number:	8G4
Expiry date:	2021-03-01
Number of tests:	12
Number of wells per test:	6+1

CHANGES COMPARED TO THE PREVIOUS HLA-B*07 ADD-ON LOT (4E4):

Well	5'-primer	3'-primer	rationale
4	Exchanged	Exchanged	Primer pair exchanged for improved HLA-specific amplification.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

The HLA-B*07:02, 07:05, 07:06, 07:61 and 07:161N alleles give different patterns in the B*07 Add-on subtyping kit¹; www.ebi.ac.uk/imgt/hla, 2018-April-16, release 3.32.0.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

RESOLUTION IN HOMO- AND HETEROZYGOTES:

Excellent.

INFLUENCE ON THE INTERPRETATION OF B*07 ADD-ON SUBTYPINGS BY NON- HLA-B*07 ALLELES:

None of importance.