

101.551-06 – including *Taq* polymerase

101.551-06u – without *Taq* polymerase

Lot No.: **1G6**

Olerup SSP® HLA-B*78

Product number: 101.551-06 – including *Taq* pol.
101.551-06u – without *Taq* pol.

Lot number: 1G6

Expiry date: 2022-03-01

Number of tests: 6

Number of tubes per test: 15+1

CHANGES COMPARED TO THE PREVIOUS HLA-B*78 LOT (5F5):

Well	5'-primer	3'-primer	rationale
9	-	Modified	3'-primer modified for increased amplification.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

B*78:01 to B*78:09, i.e. all the currently recognized HLA-B*78 alleles, give rise to unique amplification patterns¹; www.ebi.ac.uk/imgt/hla, 2017-October-27, release 3.30.0.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

RESOLUTION IN HLA-B*78 HOMO- AND HETEROZYGOTES:

The B*78:01,78:01, B*78:01,78:02 and B*78:02,78:02 genotypes give rise to unique amplification patterns.

INFLUENCE ON THE INTERPRETATION OF HLA-B*78 SUBTYPINGS BY NON-HLA-B*78 ALLELES:

None frequently occurring.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No suggestions received.

Changes in revision R01 compared to R00:

1. The expiration date has been altered due to extension of shelf-life.