

HLA-B*82
101.552-06 – including *Taq* polymerase
101.552-06u – without *Taq* polymerase
Lot No.: 9H6

Release Note

Page 1 of 1

Olerup SSP[®] HLA-B*82

| | |
|----------------------------------|---|
| Product number: | 101.552-06 – including <i>Taq</i> pol. 101.552-06u – without <i>Taq</i> pol. |
| Lot number: | 9H6 |
| Expiry date: | 2023-08-01 |
| Number of tests: | 6 |
| Number of wells per test: | 5+1 |

CHANGES COMPARED TO THE PREVIOUS HLA-B*82 LOT (4H2):

The HLA-B*82 primer set is unchanged compared to the previous *Olerup SSP[®]* HLA-B*82.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

B*82:01 to B*82:03, i.e. all the currently recognized HLA-B*82 alleles, give rise to unique amplification patterns; www.ebi.ac.uk/imgt/hla, 2019-January-23, release 3.35.0¹.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

RESOLUTION IN HLA-B*82 HOMO- AND HETEROZYGOTES:

Excellent.

INFLUENCE ON THE INTERPRETATION OF HLA-B*82 SUBTYPINGS BY NON-HLA-B*82 ALLELES:

None of importance.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.

Changes in revision R01 compared to R00:

1. The expiration date has been altered due to shelf-life extension.



0197

For *In Vitro* Diagnostic Use
 MA125 v01 SSP Release Note Template
 Date August 2021, Rev. No: 01