

HLA-B*82 Release Note Page 1 of 1

101.552-06 – including *Taq* polymerase 101.552-06u – without *Taq* polymerase

Lot No.: 2R7

Olerup SSP® HLA-B*82

Product number: 101.552-06 – including *Taq* pol.

101.552-06u - without *Taq* pol.

Lot number: 2R7

Expiry date: 2026-09-01

Number of tests: 6 Number of wells per test: 5+1

CHANGES COMPARED TO THE PREVIOUS HLA-B*82 LOT (9H6):

The HLA-B*82 primer set is unchanged compared to the previous lot.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

B*82:01 to B*82:04, i.e. all the currently recognized HLA-B*82 alleles, will be amplified by the primers in the HLA-B*82 subtyping kit^{1,2}; www.ebi.ac.uk/imgt/hla, 2022-July-12, release 3.49.0.

The HLA-B*82 kit enables separation of the confirmed HLA-B*82 alleles as listed in the IMGT/HLA database 3.35.0. An HLA allele is listed as confirmed by IMGT/HLA if it has been sequenced by more than a single laboratory or from multiple sources.

The HLA-B*82 kit also enables identification of many null and alternatively expressed alleles.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page http://hla.alleles.org/alleles/deleted.html.

RESOLUTION IN HLA-B*82 HOMO- AND HETEROZYGOTES:

Excellent.

INFLUENCE ON THE INTERPRETATION OF HLA-B*82 SUBTYPINGS BY NON-HLA-B*82 ALLELES:

None of importance.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.



For *In Vitro* Diagnostic Use MA125 v01 SSP Release Note Template Date: October 2022, Rev. No: 00