Olerup SSP[®] HLA-C*04 Add-on

Product number:	101.862-12 – including <i>Taq</i> polymerase
	101.862-12u – without <i>Taq</i> polymerase
Lot number:	0H8
Expiry date:	2021-04-01
Number of tests:	12
Number of wells per test:	4+1

The HLA-C*04 Add-on primer set is unchanged compared to the previous *Olerup* SSP[®] HLA-C*04 Add-on (Lot No. 6E1).

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

The HLA-C*04:01, 04:09N, 04:30 and 04:82 alleles give different patterns in the HLA-C*04 Add-on kit¹, <u>www.ebi.ac.uk/imgt/hla</u>, 2018-July-11, release 3.33.0.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <u>http://hla.alleles.org/alleles/deleted.html</u>.

RESOLUTION IN HOMO- AND HETEROZYGOTES: Good.

INFLUENCE ON THE INTERPRETATION OF HLA-C*04 ADD-ON SUBTYPINGS BY NON-HLA-C*04 ALLELES:

None frequently occurring.