

## **Olerup SSP<sup>®</sup> HLA-C\*07**

**Product number:** 101.615-24/04 – including *Taq* pol.  
101.615-24u/04u - without *Taq* pol.  
**Lot number:** 03M  
**Expiry date:** 2013-October-01  
**Number of tests:** 24 tests – Product No. 101.615-24/24u  
4 tests – Product No. 101.615-04/04u  
**Number of wells per test:** 48

### **CHANGES COMPARED TO THE PREVIOUS HLA-C\*07 LOT (04H):**

Well	5'-primer	3'-primer	rationale
23	Modified	-	Increased yield of specific PCR product.
33	Modified	-	Increased yield of specific PCR product.

**THE NUMBER OF WELLS** is unchanged.

### **ALLELE COVERAGE:**

C\*07:01 to C\*07:99, i.e. all the currently recognized HLA-C\*07 alleles, give rise to unique amplification patterns; [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla), 2010-April-01, release 3.0.0.

### **RESOLUTION IN HLA-C\*07 HOMO- AND HETEROZYGOTES:**

The C\*07:01, 07:01, C\*07:01,07:02 and C\*07:02,07:02 genotypes give rise to unique amplification patterns.

### **INFLUENCE ON THE INTERPRETATION OF HLA-C\*07 SUBTYPINGS BY NON-HLA-C\*07 ALLELES:**

None frequently occurring.

### **MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:**

No comments received.