

101.863-12 – including *Taq* polymerase101.863-12u – without *Taq* polymeraseLot No.: **0F6****Olerup SSP® HLA-C*07 Add-on**

Product number: 101.863-12 – including *Taq* polymerase
101.863-12u – without *Taq* polymerase

Lot number: 0F6

Expiry date: 2019-08-01

Number of tests: 12

Number of wells per test: 7+1

The HLA-C*07 Add-on primer set is unchanged compared to the previous *Olerup SSP® HLA-C*07 Add-on (Lot No. 93Y)*.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

The HLA-C*07:06, C*07:18, C*07:343 and C*07:419 and the C*07:01, the C*07:50 and the C*07:349 and the C*07:02 and the C*07:11 and the C*07:04 alleles are uniquely separated from each other in the HLA-C*07 Add-on kit, www.ebi.ac.uk/imgt/hla, 2016-October-14, release 3.26.0¹.

The HLA-C*07 Add-on kit cannot distinguish the silent mutations in the C*07:01:01:01-07:01:58 alleles, the C*07:02:01:01-07:02:77 alleles or the C*07:04:01:01-C*07:04:11 alleles.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

RESOLUTION IN HOMO- AND HETEROZYGOTES:

Good.

INFLUENCE ON THE INTERPRETATION OF HLA-C*07 ADD-ON SUBTYPINGS BY NON-HLA-C*07 ALLELES:

None frequently occurring.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.