

Olerup SSP[®] HLA-C*07 Add-on

Product number:	101.863-12 – including <i>Taq</i> polymerase 101.863-12u – without <i>Taq</i> polymerase
Lot number:	2H3
Expiry date:	2022-12-01
Number of tests:	12
Number of wells per test:	7+1

CHANGES COMPARED TO THE PREVIOUS HLA-C*07 ADD-ON LOT (0F6):

The HLA-C*07 Add-on primer set is unchanged compared to the previous *Olerup SSP[®] HLA-C*07 Add-on (Lot No. 0F6)*.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

The HLA-C*07:06, C*07:18, C*07:343 and C*07:419 and the C*07:01, the C*07:50 and the C*07:349 and the C*07:02 and the C*07:11 and the C*07:04 alleles are uniquely separated from each other in the HLA-C*07 Add-on kit¹, www.ebi.ac.uk/imgt/hla, 2018-July-11, release 3.33.0.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

RESOLUTION IN HOMO- AND HETEROZYGOTES:

Good.

INFLUENCE ON THE INTERPRETATION OF HLA-C*07 ADD-ON SUBTYPINGS BY NON-HLA-C*07 ALLELES:

None frequently occurring.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.

Changes in revision R01 compared to R00:

1. The expiration date has been altered due to shelf-life extension.