

101.865-12 – including *Taq* polymerase

101.865-12u – without *Taq* polymerase

Lot No.: **0H6**

Olerup SSP® HLA-C*15 Add-on

Product number:	101.865-12 – including <i>Taq</i> polymerase 101.865-12u – without <i>Taq</i> polymerase
Lot number:	0H6
Expiry date:	2022-10-01
Number of tests:	12
Number of wells per test:	3+1

CHANGES COMPARED TO THE PREVIOUS HLA-C*15 ADD-ON LOT (5E7):

The HLA-C*15 Add-on primer set is unchanged compared to the previous *Olerup SSP®* HLA-C*15 Add-on (Lot No. 5E7).

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

The HLA-C*15:02, 15:05, 15:29 and 15:87 alleles give different patterns in the HLA-C*15 Add-on kit¹; www.ebi.ac.uk/imgt/hla, 2018-July-11, release 3.33.0.

¹Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

RESOLUTION IN HOMO- AND HETEROZYGOTES:

Good.

INFLUENCE ON THE INTERPRETATION OF HLA-C*15 ADD-ON SUBTYPINGS BY NON-HLA-C*15 ALLELES:

None.

Changes in revision R01 compared to R00:

1. The expiration date has been altered due to shelf-life extension.