

**101.628-06 – including *Taq* polymerase**

**101.628-06u – without *Taq* polymerase**

**Lot No.: 2G9**

## **Olerup SSP® HLA-C\*17**

<b>Product number:</b>	<b>101.628-06 – including <i>Taq</i> polymerase</b> <b>101.628-06u – without <i>Taq</i> polymerase</b>
<b>Lot number:</b>	<b>2G9</b>
<b>Expiry date:</b>	<b>2021-01-01</b>
<b>Number of tests:</b>	<b>6</b>
<b>Number of wells per test:</b>	<b>15+1</b>

### **CHANGES COMPARED TO THE PREVIOUS HLA-C\*17 LOT (1E9):**

<b>Well</b>	<b>5'-primer</b>	<b>3'-primer</b>	<b>rationale</b>
4	-	Removed	Excess 3'-primer removed.

**THE NUMBER OF WELLS** is unchanged.

### **ALLELE COVERAGE:**

C\*17:01 to C\*17:38, i.e. all the currently recognized HLA-C\*08 alleles, will be amplified by the primers in the HLA-C\*08 SSP kit<sup>1</sup>; [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla), 2018-January-19, release 3.31.0.

The HLA-C\*17 kit enables separation of the confirmed HLA-C\*17 alleles as listed in the IMGT/HLA database 3.24. An HLA allele is listed as confirmed by IMGT/HLA if it has been sequenced by more than a single laboratory or from multiple sources.

The HLA-C\*17 kit also enables identification of null and alternatively expressed alleles.

<sup>1</sup>Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

### **RESOLUTION IN HLA-C\*17 HOMO- AND HETEROZYGOTES:**

Excellent.

### **INFLUENCE ON THE INTERPRETATION OF HLA-C\*17 SUBTYPINGS BY NON-HLA-C\*17 ALLELES:**

Non of importance.

### **MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:**

No comments received.