

Olerup SSP® HLA-Cw*01

Product number: 101.621-12 – including *Taq* polymerase
101.621-12u – without *Taq* polymerase
Lot number: 44E
Expiry date: 2010-April-01
Number of tests: 12
Number of wells per test: 16

CHANGES COMPARED TO THE PREVIOUS HLA-Cw*01 LOT (X34):

Well	5'-primer	3'-primer	rationale
2	-	Added	Primer added for the Cw*0115 allele.
3	Added	Added	Primer pair added for the Cw*0114 allele.
4	Added	Added	Primer pair added for the Cw*0116 allele.
5	-	Added	Primer added for the Cw*0120 allele.
7	-	Added	Primer added for the Cw*0120 allele.
8	Modified	-	Increased specificity of specific primer pair.
14	New	New	New primer pair for the Cw*0117 allele.
15	New	New	New primer pair for the Cw*0118 allele.
16	New	New	New primer pair for the Cw*0119 allele.

THE NUMBER OF WELLS has been increased from 13 to 16.

ALLELE COVERAGE:

Cw*0102 to Cw*0120 i.e. all the currently recognized HLA-Cw*01 alleles, give rise to unique amplification patterns; www.ebi.ac.uk/imgt/hla, 2008-April-08, release 2.21.0.

RESOLUTION IN HLA-Cw*01 HOMO- AND HETEROZYGOTES:

The Cw*0102,0102 genotype give rise to a unique amplification pattern.

INFLUENCE ON THE INTERPRETATION OF HLA-Cw*01 SUBTYPINGS BY NON-HLA-Cw*01 ALLELES:

None.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.