HLA-Cw*08 Release Note Page 1 of 1

101.623-12 – including *Taq* polymerase 101.623-12u – without *Taq* polymerase

Lot No.: **79F**

Olerup SSP® HLA-Cw*08

Product number: 101.623-12 – with *Taq* polymerase

101.623-12u – without *Taq* polymerase

Lot number: 79F

Expiry date: 2011-May-01

Number of tests: 12 Number of wells per test: 16

CHANGES COMPARED TO THE PREVIOUS HLA-CW*08 LOT (23E):

Well	5'-primer	3'-primer	rationale
6	-	Added	New primer added for the Cw*0823 allele.
9	Exchanged	Exchanged	New primer pair for improved resolution.
10	Added	Added	New primer pair for the Cw*0817 allele.
11	Added	Added	New primer pair for the Cw*0820 allele.
14	Added	Added	New primer pair for the Cw*0818 allele.
15	Added	Added	New primer pair for the Cw*0819 allele.
16	Added	Added	New primer pair for the Cw*0822 allele.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

Cw*0801 to Cw*0823, i.e. all the currently recognized HLA-Cw*08 alleles, give rise to unique amplification patterns; www.ebi.ac.uk/imgt/hla, 2009-May-09, release 2.25.2.

RESOLUTION IN HLA-CW*08 HOMO- AND HETEROZYGOTES:

The Cw*0801, 0801 and the Cw*0801, 0816 genotypes give rise to identical amplification pattern.

INFLUENCE ON THE INTERPRETATION OF HLA-CW*08 SUBTYPINGS BY NON-HLA-CW*08 ALLELES:

None frequently occurring.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.