

Lot No.: **04E**

## Olerup SSP<sup>®</sup> HLA-Cw\*12

**Product number:** 101.624-12 – including *Taq* polymerase  
101.624-12u – without *Taq* polymerase  
**Lot number:** 04E  
**Expiry date:** 2010-January-01  
**Number of tests:** 12  
**Number of wells per test:** 22

### CHANGES COMPARED TO THE PREVIOUS HLA-Cw\*12 LOT (X10):

Tube	5'-primer	3'-primer	rationale
21	New	New	New primer pair for the Cw*1220 allele.
22	New	New	New primer pair for the Cw*1221 allele

**THE NUMBER OF WELLS** has been increased from 20 to 22.

#### ALLELE COVERAGE:

Cw\*1202 to Cw\*1221<sup>1</sup>, i.e. all the currently recognized HLA-Cw\*12 alleles, give rise to unique amplification patterns; [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla), 2007-October-05, release 2.19.0.

<sup>1</sup>The HLA-Cw\*12 primer set cannot separate the Cw\*1209 and Cw\*0516 alleles. These two alleles can be distinguished by the HLA-Cw low resolution and/or HLA-Cw\*05 kits.

#### RESOLUTION IN HLA-Cw\*12 HOMO- AND HETEROZYGOTES:

Good.

#### INFLUENCE ON THE INTERPRETATION OF HLA-Cw\*12 SUBTYPINGS BY NON-HLA-Cw\*12 ALLELES:

None frequently occurring.

#### MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.