# *Olerup* SSP<sup>®</sup> HLA-Cw\*16

Product number:	101.627-12 – including <i>Taq</i> polymerase
	101.627-12u – without <i>Taq</i> polymerase
Lot number:	22E
Expiry date:	2010-March-01
Number of tests:	12
Number of wells per test:	8

## CHANGES COMPARED TO THE PREVIOUS HLA-Cw\*16 LOT (LX29):

Well	5'-primer	3'-primer	rationale
6	Added	-	Primer added for the Cw*1610 allele.

THE NUMBER OF WELLS is unchanged.

## ALLELE COVERAGE:

Cw\*1601 to Cw\*1610 i.e. all the currently recognized HLA-Cw\*16 alleles, give rise to unique amplification patterns; <u>www.ebi.ac.uk/imgt/hla</u>, 2008-January-11, release 2.20.0.

#### **RESOLUTION IN HLA-CW\*16 HOMO- AND HETEROZYGOTES:** Good.

# Influence on the interpretation of HLA-Cw\*16 subtypings by non-HLA-Cw\*16 alleles:

None frequently occurring.

# MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No comments received.